



बुङ्गल नगरपालिका  
नगर कार्यपालिकाको कार्यालय  
खिरातडी, बझाङ्ग

प.सं.: ०८२।०८३

च.नं.: ८८९



☎ 9766973710 कार्यालय प्रमुख  
URL: [www.bungalmun.gov.np](http://www.bungalmun.gov.np)  
e-mail: [info@bungalmun.gov.np](mailto:info@bungalmun.gov.np)  
e-mail: [ito.bungalmun@gmail.com](mailto:ito.bungalmun@gmail.com)  
e-mail: [ito.bungalmun.gov.np](mailto:ito.bungalmun.gov.np)  
सुदूरपश्चिम प्रदेश नेपाल

मिति: २०८२।१२।०१


नं.सं. ११४६ चिल्लागा १० आईतवार

बिषय: वजार दर भाउ पठाई दिने वारे ।

श्री सम्बन्धित सबै ।

प्रस्तुत बिषय सम्बन्धमा यस बुङ्गल नगरपालिकाको आ.व. २०८२।०८३ को वार्षिक स्वीकृत कार्यक्रम अनुसार तपशिल बमोजिमका स्वास्थ्य औजार उपकरणहरु खरिद गर्नु पर्ने भएकाले उक्त उपकरणहरुको यसै पत्र साथ संलग्न टेक्निकल स्पेसिफिकेशनको आधारमा पत्र प्राप्त मिति ले ७ दिन भित्र सामाग्रीहरुको वजार दर भाउ उपलब्ध गरिदिनुहुन सम्बन्धीत सबैमा अनुरोध गरिन्छ ।

  
कलठ बहादुर बौहरा  
नि.स्वास्थ्य शाखा प्रमुख

  
अम्बर राज जोशी  
प्रमुख प्रशासकिय अधिकृत  
प्रमुख प्रशासकीय अधिकृत

तपशिल:

१. USG Portable Colour Doppler — थान १
२. ECG Machine, Portable (१२ channel) — थान १
३. ५००mA High Frequency X-Ray Machine with DR system — थान १
४. Post Mortem Set — थान १
५. Refrigerator, Ice Lined — थान ३
६. Portable USG — थान १

"व्यवसायिक र सिर्जनशील प्रशासन:-विकास समृद्धि र सुशासन"

1. Technical specification for refrigerator

- Refrigerator Temperature ;2-8 deg. C , Capacity; 120 liter or more
- Temperature Display Board & Alarm System.
- Door with locking mechanism
- Voltage Stabilizer.
- 2Year Warranty.
- ISO, CE or USFDA

2. Short Technical specification for x-ray with DR and Printer


- High Frequency 500mA x ray with Three phase voltage supply
- Output Voltage -40kw or more
- Microprocessor based high frequency.
- APR program.
- Control Panel display size 7 inch or more
- Flat panel Detector at least 35 x 43 cm with pixel size 150 um or less and weight less than 350.g
- Double tray imaging printer with at least 50 sheets per hour of size 14 x17 inch.
- Dr panel, printer and workstation software from same brand or company
- X-ray ISO, CE or USFDA CERTIFIED
- DR panel ISO, CE & USFDA
- Printer ISO, CE or USFDA Certified
- Warranty 2 years.

3. Short Technical Specification of ECG Machine:

- Portable digital ECG machine ( 12 channel).
- Simultaneous recording of 12 standard leads: aVR, aVL, aVF, I, II, III and VI-6 pre-cordials.
- Internal memory for data storage.
- Alphanumeric keyboard with function keys.
- Alphanumeric LCD display, Minimum 7 inches display size or more.
- Number of channels printed is user selectable: 3, 6 or 12.
- Paper speed, user adjustable: 5, 25 and 50mm/sec.
- With internal re-chargeable battery
- Certification: ISO, CE or USFDA
- Warranty 2 years.

  
Er. Triyogi Chaudhary  
Bio-Medical Engineer 'A'  
NEC Reg. No. 226



  
प्रमुख प्रशासकीय अधिकृत

#### 4. Short Technical Specification of USG With echo


- High end USG machine with 2 probes.
- Upgradable to 3D & 4D feature.
- High resolution of 1920 x 1080 or better with display size 15.6 inch or more
- Additional 10 inch or more touch screen monitor
- Cardiac measurement MVA & AVA
- Features like Tissue Doppler imaging, needle guide tips, Auto IMT, Auto NT, Auto OB measurement.
- DICOM © 3.0 Connectivity
- ISO, CE and USFDA approved
- Warranty 2 years.

#### 5. Short Technical specification for Tablet Ultrasound.

- portable ultrasound scanner with convex probe
- Multiple imaging modes B-mode, C-mode, M-mode and pulseDoppler
- Wt. less than 3kg or less
- Auto-Optimization Features: Auto Trace, Auto BLineand one-touch optimization for B/C/PW
- Modes & eVocal Voice Assistant.
- Warranty 2 years.
- ISO, CE and USFDA approved.

  
Er. Triyogj Chaudhary  
Bio-Medical Engineer 'A'  
NEC Reg. No. 226



  
प्रमुख प्रशासकीय अधिकृत

## 6. Specification for Autopsy/ post mortem examination kit

- Set typically features scalpels, scissors, forceps, needles, and a variety of specialized instruments.
- Autopsy set must have 14 essential tools or more
- Material: stainless steel
- Steel handle
- Durable box

*Er. Triyugi Chaudhary*

Er. Triyugi Chaudhary  
Bio-Medical Engineer 'A'  
NEC Reg. No. 226

*3/11/21*  
प्रमुख प्रशासकीय अधिकृत

